

# BIOCOMPATIBILITY CERTIFICATE

**Testmaterial:** Elastosil® LR3003/20 A,B Control No: 1823SL

**Manufacturer:** WACKER Chemie GmbH, Germany  
WACKER Silicones Corporation, USA  
WACKER Chemicals East Asia Ltd., Japan

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the test item according to ISO 10993-1:

**CYTOTOXICITY**

**HEMOLYSIS**

**PYROGENICITY**

**SENSITISATION**

**DERMAL IRRITATION**

**IMPLANTATION (90 days)**


**Results:** The test item showed no relevant biological effects in the studies performed.

Suitable for use in manufacturing of medical devices for short term (no longer than 30 days) implant applications in accordance with the Wacker Health Care Letter and Guidelines.

**BSL BIOSERVICE Scientific Laboratories GmbH Munich**

Behringstraße 6

D-82152 Planegg



Dr. Albrecht Poth

Manager Biological Safety Testing



Date: December 28, 1998